



深圳华信计量检测科技有限公司

ShenZhen Huaxin Metrology and Detection Technology Co.,Ltd.



中国认可  
国际互认  
校准  
CALIBRATION  
CNAS L10016

# 校准证书

## CALIBRATION CERTIFICATE

证书编号:

Certificate No.

委托方

Client

委托方地址

Address

仪器名称

Description

型号/规格

Model/Type

制造厂

Manufacturer

出厂编号

Serial No.

管理编号

Asset No.

结论

Conclusion

Wedge probe

IEC 62368-1 Figure V.4

Kingpo Technology Development Limited

接收日期

Receive Date

校准日期

Calibration Date

建议复校日期

Suggested Recal.Date

2020-12-21

2020-12-21

2021-12-20

KP0201210502

/

所校准项目合格

批准

Approved by

吕波

吕波

职务

Position

技术负责人

核验

Inspected by

郭芳倩

郭芳倩

校准

Calibrated by

李星虎

李星虎



计量校准机构备案证书号 (Register No.): 粤校备2017B016

地址: 广东省深圳市宝安区燕罗街道塘下涌社区致和路12号广盛大厦201

Address: Room 201, Guangsheng Building, No.12,Zhihe Road, Tangxiayong Community,Yanluo Street, Bao'an District, Shenzhen,Guangdong, China Street, Bao'an District, Shenzhen,Guangdong, China

电话Tel: 0755-8525 9375

电子邮件Email: hxjlc@yeah.net

传真Fax: 0755-8525 9376

邮编Post Code: 518035



# 校准结果

证书编号:HX2012CC228002  
Certificate No.

## RESULTS OF CALIBRATION

### 一、外观及各工作性能检查: (Appearance & Performance check)

- 1.1.外观: (Appearance) 符合要求 (Pass)  
1.2.工作性能检查: (Performance check) 正常 (Pass)

### 三、尺寸测量: (Size measurement)

代码 code	标准要求(mm) Standard requirement	实测值(mm) measured value	$U$ (mm) $k=2$	结论 (P/F) conclusion
1	300.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	300.083	0.01	P
2	180.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	179.747	0.01	P
3	12.00 $\begin{matrix} +0.13 \\ -0.13 \end{matrix}$	12.101	0.01	P
4	50.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	49.987	0.01	P
5	30.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	30.040	0.01	P
6	60.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	59.960	0.01	P
7	100.00 $\begin{matrix} +0.3 \\ -0.3 \end{matrix}$	100.143	0.01	P
8	2.00 $\begin{matrix} +0.13 \\ -0.13 \end{matrix}$	2.014	0.01	P
9	4.00 $\begin{matrix} +0.13 \\ -0.13 \end{matrix}$	3.988	0.01	P
10	24.00 $\begin{matrix} +0.13 \\ -0.13 \end{matrix}$	24.020	0.01	P

